



CERTIFICATE OF CONFORMITY

Order code: **222-055**

Reference Thermometer

Range: -199.99 to 199.99°C

Resolution: 0.01°C

Accuracy: $\pm 0.03^\circ\text{C}$: -49.99 to 149.99°C, $\pm 0.1^\circ\text{C}$: -150 to 200°C

Made in Britain

Conforms to RoHS & REACH Legislation

Also conforms to the following regulations; EC Reg 1935/2004, EU Reg 10/2011 & amendment 2020/1245, BS EN13485, Directive 2011/65/EU & amendment 2015/863, EU Reg 2014/35/EU, EU Reg 2014/30/EU.

DECLARATION OF CONFORMITY

We certify that the goods shown on this certificate are manufactured, tested, inspected and stored, to conform in all respects to the stated specifications and in accordance with the Quality System of ETI Limited, accredited to BS EN ISO 9001, as approved and audited by the British Standards Institution.

We also confirm that when used in an appropriate manner, the products are suitable for use within the Food Industry. Where appropriate and required, the product complies with EEC Directives with regard to European Legislation and UKCA/CE Marking requirements. Including EC Regulation No 1935/2004 on Materials and articles intended to come into contact with food & Commission Regulation EU Reg 10/2011 & amendment 2020/1245 on plastic materials and articles intended to come into contact with food. The plastic material of this product contains biomaster additive in the plastic mixture. This product conforms to EN 13485, Both, S, E, 1, 0.5, -150 °C +200°C.

All ETI supplied instruments are suitable for use within HACCP plans where the quoted specification for critical control points match the performance of the instrument.

Signed, on behalf of Electronic Temperature Instruments Ltd

A handwritten signature in black ink, appearing to read 'Andy Reid', is positioned above the printed name.

Andy Reid

Quality Manager

Valid as of date of purchase.

Easting Close
Worthing
West Sussex
BN14 8HQ
T +44 (0)1903 202151
E sales@etiltd.com
W etiltd.com



SERTIFIKAT O USAGLAŠENOSTI

Artikal broj : **222-055**

Referentni termometar

Opseg : -199.99 to 199.99°C

Rezolucija : 0.01°C

Tačnost: $\pm 0.03^\circ\text{C}$: -49.99 to 149.99°C, $\pm 0.1^\circ\text{C}$: -150 to 200°C

Proizvedeno u Britaniji

U skladu sa RoHS & REACH zakonodavstvom

Takođe je u skladu sa sledećim propisima; EC Reg 1935/2004, EU Reg 10/2011 i amandman 2020/1245, BS EN13485, Directive 2011/65/EU i amandman 2015/863, EU Reg 2014/35/EU, EU Reg 03/EU4.

IZJAVA O USAGLAŠENOSTI

Potvrđujemo da je roba prikazana na ovom sertifikatu proizvedena, testirana, pregledana i uskladištena, u skladu sa svim aspektima navedenim specifikacijama i u skladu sa Sistemom kvaliteta kompanije ETI Limited, akreditovanom prema BS EN ISO 9001, kako je odobrila i proveravala britanske institucije za standarde.

Takođe potvrđujemo da kada se koriste na odgovarajući način, proizvodi su pogodni za upotrebu u prehrambenoj industriji. Tamo gde je prikladno i potrebno, proizvod je usklađen sa direktivama EEZ u vezi sa evropskim zakonodavstvom i zahtevima UKCA/CE oznake. Uključujući Uredbu EC br. 1935/2004 o materijalima i predmetima namenjenim da dođu u kontakt sa hranom i Uredbu Komisije EU Reg 10/2011 i amandman 2020/1245 o plastičnim materijalima i predmetima namenjenim da dođu u kontakt sa hranom. Plastični materijal ovog proizvoda sadrži biomaster aditiv u plastičnoj mešavini. Ovaj proizvod je usklađen sa EN 13485, Oba, S, E, 1, 0,5, -150 °C +200 °C.

Svi instrumenti koje isporučuje ETI pogodni su za upotrebu u okviru HACCP planova gde se navedena specifikacija za kritične kontrolne tačke poklapa sa performansama instrumenta.

Potpisano, u ime Electronic Temperature Instruments Ltd

Andy Reid

Quality Manager

Valid as of date of purchase.

Easting Close
Worthing
West Sussex
BN14 8HQ
T +44 (0)1903 202151
E sales@etiltd.com
W etiltd.com

ROVEX Inženjering doo

Bulevar Crvene Armije 11b/lok7;
11070 Beograd;
tel: 011-3115208;
e-mail:prodaja@rovex.co.rs